

NASCO LF01199 AUSCULTATION POSTERIOR BOARD INSTRUCTION SHEET

Attaching to a Case Without Side Rails

If you have a case without rails on the inside of the cover, with the metal Serial Tag riveted in the middle, and with the three scope pockets, follow these instructions:

1. Make note of the Serial No. on the backside of the board and on the inside case cover for your records and future use. Doing this now will prevent you from possibly having to remove the board to obtain the Serial No. in the future.
 2. There are four 1" square, double-sided pieces of tape enclosed, which are to be used to attach the auscultation board to your case. The tape squares can be utilized in any manner you wish, but the following is the recommended method:
 - a. Remove the backing from the tape pieces.
 - b. Fold each tape piece in half – over onto itself.
 - c. Place each tape piece approximately 2" in from the corners on the backside of the board.
 - d. Center and firmly press the board onto the case cover.
-

Attaching to a Case with Side Rails

If your case has side rails for holding the board onto the case, follow these instructions:

1. Carefully slide the board beneath the two side rails that are on the inner sidewalls of the case. Push it all the way in so that the bottom of the board slips beneath the bottom rail.
2. There are four 1" square, double-sided pieces of tape enclosed, which are to be used to keep the auscultation board attached to the case and prevent it from sliding out. You may use the tape squares in any manner you wish, but the following is the recommended method:
 - a. Remove the backing from the tape pieces.
 - b. Double stack the four tape squares so that you have two stacks of doubled tape.
 - c. While lifting the top edge of the board away from the case surface, insert each of the two tape stacks between the board and the case surfaces – adhere to one of the surfaces. They should be equidistant on the width of the board and about 1" down from the top board.
 - d. Let go of the board so that the tape stacks come into contact with the other surface, and press firmly to achieve a strong bond.

Nasco **Life/form**®

**Auscultation Trainer and
SmartScope™
LF01142U & LF01172U
Instruction Manual**



***Life/form*® Products by Nasco**

About the Simulator

The **Life/form®** Auscultation Trainer Simulator features 28 heart and lung conditions selected by the instructor by wireless remote control. The simulator duplicates human conditions as closely as modern plastics and electronic technology allows.

The remote control does not have to be pointed directly at the manikin or SmartScope™ to operate. One remote control will operate multiple sets of SmartScopes™ and manikins simultaneously. Great for group instruction. The range of the remote control is up to 100 feet.

LF01172U includes amplified speaker. LF01142U does not include amplified speaker.

List of Components

- Trainer Torso
- Remote Control
- SmartScope™ with Single- & Dual-Use Headpieces
- Amplifier/Speaker System (LF01189U included in LF01172U only)
- Batteries
- Colored Dots
- Hard Carry Case

Set Up

- Remove product from packaging.
- Install batteries first in the remote and then in the SmartScope™. (**See figure 1.**)
- Retain packaging for storage.

General Instructions for Use

1. To begin using the Auscultation Trainer, press the red power button. This turns on the remote control and sends a signal to activate the SmartScope™ as well. After the unit is activated, the display will be in the “status” mode, displaying the current menu settings for the heart and lung conditions.
2. To select a new condition, press either the heart or lung button. This will put the display into the menu mode. The user can select a condition by either using the number buttons, or by viewing the conditions in sequence using the scroll button
3. After the condition is selected, press the enter button to activate.



Figure 1

The remote control can be programmed to these heart conditions:

- 01 Normal Heart
- 02 Aortic regurgitation
- 03 Pulmonary stenosis
- 04 Mitral stenosis
- 05 Holosystolic
- 06 Mid-systolic
- 07 S3 Gallop
- 08 S4 Gallop
- 09 Systolic click
- 10 ASD
- 11 PDA
- 12 VSD

The remote control can be programmed to these lung conditions:

- 01 Normal lung
- 02 Vesicular
- 03 Wheezes
- 04 Mono wheeze
- 05 Fine crackle
- 06 Coarse crackle
- 07 Ronchi crackle
- 08 Stridor
- 09 Cavernous
- 10 Bronchovesicular
- 11 Bronchial
- 12 Pulmonary edema
- 13 Infant
- 14 Friction rub
- 15 Egophony
- 16 Pectoriloquy

4. To listen to the selected sounds, place the ear pieces in ears angled in a forward position. Place the diaphragm of the SmartScope™ over one of the appropriate sites on the manikin. (See diagram on pages 4 & 5.)
5. As an alternative, an amplified speaker (LF01189U, included with LF01172U) can be used, allowing an entire classroom to hear at the same time. To connect the speaker, plug the speaker cord into the speaker jack on the top of the SmartScope™ box. (See **figure 2.**) When a speaker is connected to the SmartScope™, the ear pieces will not work. Place the diaphragm of the SmartScope™ over the appropriate sites on the manikin. (See diagram on pages 4 & 5.)
6. Colored dots are included with the torso to aid with teaching.

Care and Maintenance

- Clean torso when needed using Nasco Cleaner (LF09919U).
- Store in hard carry case in a temperature controlled environment, avoiding extreme hot or cold temperatures.
- Replace batteries regularly for optimum working condition.
- It is recommended that batteries are removed when not in use.
- Avoid getting ink on torso simulator or practice boards. Ink will transfer indelible stains.

Troubleshooting

- When preparing to use your SmartScope™ and remote, use batteries that are fully charged and installed correctly. (See **figure 3.**) The SmartScope™ is not equipped with a power button; this may drain the useful life of your batteries.
- Due to frequency changes, Adult Auscultation Trainer SmartScopes™ with serial numbers **0 to 4999** are only compatible with remotes with serial numbers **0 to 4999**. SmartScopes™ with serial numbers **5000 and up** will only work with remotes with serial numbers **5000 and up**. The serial numbers are located on the back of the device and are handwritten in ink. (See **figure 4.**)



Figure 2



Figure 3



Figure 4

Common Causes of No Sounds

- **Incorrect SmartScope™ placement.** Move the SmartScope™ slowly over the torso to locate sensors. See diagrams on page 4 for placement of sensors. You may apply color-coded dots on the heart and lung sites to assist with training. (See figure 5.)
 - **Ear tips facing the wrong way.** Verify that the ear tips of the SmartScope™ are facing forward in your ear canal.
 - **Battery replacement.** After installing new batteries, make sure to turn the remote off and back on. This allows the remote and SmartScope™ to reset.
- Note:** SmartScope™ does not turn off.
- **Battery saver mode.** The remote will go into battery saver mode if left on the same setting for 8 minutes. Sounds will not be heard if the remote is off.
 - **Using multiple remotes.** One remote will operate all SmartScopes™ within 100 ft. simultaneously as long as the serial numbers are compatible. Using multiple remotes can cause interference. (See figure 6.)
 - **Interference.** Wireless communication from another transmitter using 433 MHz may cause sound selection issues or existing sounds to stop or restart. Reset the remote by turning it off briefly, then on again and select a different Heart or Lung sound.
 - **Connecting an external speaker.** Insert the jack end of the speaker cord into the top of the SmartScope™. (See figure 2.)
 - **Using an external speaker.** When using an amplifier/speaker, no sounds will be heard through the SmartScope™ ear tips.
 - **Check volume.** Make sure the volume on the amplifier/speaker is turned up.

Actual product may vary slightly from photo. Nasco reserves the right to change product color, materials, supplies, or function as needed.

Available Supplies

- LF01144U** Additional SmartScope™
- LF01148U** Replacement Remote Control
- LF09919U** Nasco Cleaner
- SB20146U** Amplified Speaker






Figure 5




Figure 6

Anterior Heart and Lung Sites




Posterior Lung Sites

Anterior Heart Sites



	Aortic Second Right Intercostal Space	Pulmonic Second Left Intercostal Space	Erb's Point Third Left Intercostal Space	Fourth Left Intercostal Space	Lower Left Sternum	Apex
1. Normal	Normal, S2 Accentuated	Normal, S2 Accentuated	Normal, S2 Accentuated	Normal	Normal	Normal
2. Aortic Regurgitation	Ejection Sound, Loud Mid systolic & soft early diastolic murmur	Ejection sound, mild systolic murmur, early diastolic blowing murmur			Normal	Normal
3. Pulmonary Stenosis	Normal	Moderate 4th sound, harsh late peaking systolic murmur, soft late pulmonic 2nd sound	Normal	Normal	Normal	Normal
4. Mitral Stenosis	Normal	Normal	Severe held expiration, tachycardia: opening snap .03 seconds after loud 2nd sound	Constrictive Pericarditis/ knock. Inspiratory augmentation indicates a gallop of right ventricular origin	Normal	Held expiration, tachycardia: opening snap, mid diastolic & presystolic murmurs, loud 1st sound
5. Holosystolic Murmur	Normal	Normal	Normal	Normal	Patient has mitral regurgitation & frequent premature ventricular contractions. Murmur is crescendo-decrescendo with late peaking. Soft S3 in mid diastole.	
6. Midsystolic Murmur	Normal	Normal	Normal	Normal	Patient with hypertrophic cardiomyopathy has a murmur that begins after S1 and ends before S2	
7. S3 Gallop	Normal	Normal	Normal	Normal	Patient has a readily heard third heart sound. S3 occurs later in diastole than the opening snap.	
8. S4 Gallop	Normal	Normal	Normal	Normal	Patient with left ventricular hypertrophy has a fourth sound (S4) that is not heard on every cycle. The sound is presystolic about .1 second before S1.	
9. Midsystolic Click	Normal	Normal	Normal	Normal	Patient has mitral prolapse which produces a mid systolic click heard during inspiration.	
10. Atrial Septal Defect	Normal	Respiration: mid systolic murmur, fixed split 2nd, soft 3rd, breath sounds with inspiration	Normal	Respiration: mid systolic murmur, fixed split 2nd, mid diastolic murmur	Normal	Normal
11. Patent Ductus Arteriosus	Normal	Continuous murmur	Normal	Continuous murmur	Normal	Normal
12. Ventricular Septal Defect	Normal	Normal	Holosystolic murmur with late crescendo		Normal	Normal

Anterior/ Posterior Lung Sites



	Tracheal Site	First Left & Right Intercostal Sites	Upper Posterior Lung Sites	Lower Posterior Lung Sites Two Midaxillary Sites 2 Lower Anterior Sites
1. Normal Lung	Tracheal	Bronchovesicular	Normal Vesicular	Normal Vesicular
2. Normal Vesicular	Tracheal	Bronchovesicular	Normal Vesicular	Normal Vesicular
3. Wheezes	Wheeze	Wheeze	Wheeze	Wheeze Lower Volume
4. Mono Wheeze	Mono Wheeze	Mono Wheeze	Mono Wheeze	Mono Wheeze Lower Volume
5. Fine Crackle	Fine Crackle	Fine Crackle	Fine Crackle	Fine Crackle
6. Coarse Crackle	Coarse Crackle	Coarse Crackle	Coarse Crackle	Coarse Crackle
7. Ronchi	Ronchi	Ronchi	Ronchi	Ronchi
8. Stridor	Stridor	Stridor	Stridor Lower Volume	Stridor Lower Volume
9. Cavernous	Cavernous	Cavernous	Cavernous	Cavernous
10. Bronchovesicular	Tracheal	Bronchovesicular	Normal Vesicular	Normal Vesicular
11. Bronchial	Bronchial	Bronchial	Normal Vesicular	Normal Vesicular
12. Pulmonary Edema	Pulmonary Edema	Pulmonary Edema	Pulmonary Edema	Pulmonary Edema
13. Infant	Infant	Infant	Infant	Infant
14. Friction Rub	Tracheal	Bronchovesicular	Friction Rub	Friction Rub
15. Egophony	Egophony	Egophony	Egophony	Egophony
16. Pectoriloquy	Pectoriloquy	Pectoriloquy	Pectoriloquy	Pectoriloquy


Other Available **Life/form®** Simulators

LF01192U Anterior Auscultation Practice Board with Case and SmartScope™

LF01200U Deluxe Auscultation Training Station

LF01220U Adult & Infant Auscultation Training Set

LF03968U Deluxe Plus **CRiSis™** Auscultation Manikin



LF01200U



LF01220U



LF01992U

Nasco

901 Janesville Avenue, P.O. Box 901
Fort Atkinson, Wisconsin 53538-0901
1.800.558.9595
eNasco.com • Email: lifeform@eNasco.com

COPYRIGHT © NASCO 2017 PRINTED IN U.S.A. NP020-17